

EC Declaration of conformity

COMPANY:
ARCTIKO A/S

ADDRESS:
LILLEBÆLTSVEJ 90,
6715 ESBJERG, DENMARK

PRODUCT INFORMATION:
ULTRA LOW TEMPERATURE UPRIGHT FREEZERS

MODEL:
ULUF P10MV / P50MV / P65MV / P90MV

WE HEREBY DECLARE THAT THE **ABOVE MENTIONED** PRODUCTS COMPLY WITH THE FOLLOWING DIRECTIVES, NATIONAL LAWS AND HARMONIZED STANDARDS:

DIRECTIVES AND NATIONAL LAWS:

2006/42/EC
2011/65/EU
2014/30/EU
2014/35/EU

MACHINERY DIRECTIVE
ROHS DIRECTIVE
ELECTROMAGNETIC COMPATIBILITY (EMC) DIRECTIVE
LOW VOLTAGE DIRECTIVE

HARMONIZED STANDARDS:

EN ISO 9001:2015
EN ISO 13485:2016
EN 61010-2-11:2019

QUALITY MANAGEMENT SYSTEM
MEDICAL DEVICES QUALITY MANAGEMENT SYSTEM
SAFETY REQUIREMENTS FOR ELECTRICAL
EQUIPMENT FOR MEASUREMENT, CONTROL AND
LABORATORY USE - PART 2-011 SPECIAL
REQUIREMENTS FOR REFRIGERATION EQUIPMENT
SAFETY REQUIREMENTS FOR ELECTRICAL
EQUIPMENT FOR MEASUREMENT, CONTROL, AND
LABORATORY USE PART 1: GENERAL REQUIREMENTS

EN 61010-1: 2010+A1

EMISSION:

EN 55011:2016 / A1:2017

EN 61000-3-3:2013 - FLICKER
EN 61000-3-2:2018 - HARMONIC CURRENT
EN 61326-1ED.3.0:2020

INDUSTRIAL; SCIENTIFIC AND MEDICAL
EQUIPMENT (ISM) - CHARACTERISTICS OF
RADIO NOISE EMISSIONS - LIMIT VALUES AND
MEASUREMENT METHODS
PRODUCT FAMILY STANDARD
PRODUCT FAMILY STANDARD
ELECTRICAL EQUIPMENT FOR MEASUREMENT,
CONTROL AND LABORATORY USE - EMC
REQUIREMENTS - PART 1: GENERAL REQUIREMENT

IMMUNITY:


IEC 61326-1ED.3.0:2020
IEC 61326-1ED.2.0:2012-07

ELECTRICAL EQUIPMENT FOR MEASUREMENT,
CONTROL, AND LABORATORY USE EMC
REQUIREMENTS PART 1: GENERAL REQUIREMENTS

Esbjerg 25/09/2024

City & DD/MM/YYYY

5082189-03



Name & position

Tove Tækker Svendsen

R&D & QA Manager